

Republic of the Philippines
Province of Davao del Norte
 Government Center, Mankilam, Tagum City

PR Number : 1301-0015

Bid Number : B20130004

Date : January 14, 2013

Page : 1

Item No.	Qty./Unit	Item	Bid Price	
			Unit Price	Total Amount

1 1 unit Defibrillator with Monitor

Power Supply Input: 220-240 VAC, 60 Hz with built in rechargeable battery pack
 Output Energy: 1 to 360 joules
 Discharge Modes: Synchronous/ Asynchronous

Features:

*Charge control buttons on control panel and on APEX paddle, Charging indicators and Isolated ECG monitoring via defibrillator

Display Parameters:

*ECG, Heart Rate, ECG Lead Selections, 1mV calibration and ECG Size

Recorder Parameters:

*Thermal Array and Speed of 25 mm/sec, with annotations

Safety Parameters:

*Alarm Indicators (Audible and Visual), Battery Condition:Charging, Low Battery, Hi and Low Heart Rate and Lead contact failure

Standard Accessories:

*Set of External Paddles (Adult and Pedia), ECG patient cables, Ground cables, Power cord, Dust cover, Accessory bag, 1 tube Gel and 1 roll/z-fold recording paper

Optional Accessories:

*Mobile cart with brake and Internal paddles

Optional Parameters:

Medical , Dental, Laboratory Equipment

APPROVED BUDGET FOR THE CONTRACT (ABC) : P 550,000.00

PEEDO-Davao del Norte Hospital IGACOS Zone

Period of Delivery : _____ day/s

**SUPPLIERS MUST SPECIFY/INDICATE
 BRAND NAMES UPON QUOTATION**

INSTRUCTIONS IN FILLING UP THIS BID FORM

Bids Distributed To:

.....

1. Bidder must read carefully the statement at the back hereof.
2. Bid Quotation must be typewritten or printed in blue or black ink.
3. Unit price, total amount and grand/lot price must be properly indicated.
4. Delivery period must be indicated in the Bid form unless otherwise provided by the procuring unit/office. Failure to indicate the period of delivery shall automatically mean 10 days delivery after receipt of the Purchase Order.
5. Place of Delivery shall be at the PGSO Warehouse unless otherwise indicated in the bid form.
6. Submit Bid in the sealed envelope.
7. Unprinted name of establishment, unprinted name of the bidder and unsigned bid shall be disqualified.
8. Erasures/alterations should be initialed accordingly.

.....
 Name of Establishment

.....
 Name of Bidder

.....
 Signature of Bidder

Republic of the Philippines
Province of Davao del Norte
 Government Center, Mankilam, Tagum City

PR Number : 1301-0015

Bid Number : B20130004

Date : January 14, 2013

Page : 2

Item No.	Qty./Unit	Item	Bid Price	
			Unit Price	Total Amount

*Pacemaker impulse rejection

Standard Requirements:

*Manuals in English Language, Operation Manual and Service Manual, Training Program for end-user on equipment's operation, One year warranty on parts and service after commissioning and acceptance, IEC/PNS compliant or its equivalent and availability of technician in Mindanao

2 1 unit Patient Monitor

Power Supply Input: 220-240 VAC , 60 Hz with built-in rechargeable battery pack

Display Parameter:

- * ECG (dual vector display)
- * Heart Rate
- * Temperature
- * SpO2
- * NIBP - with automatic measurement cycles

Standard Features:

- * Alarms for ECG, RR, HR, NIBP, SP02
- * Temperature (Audible and Visual)
- * With built in thermal printer

Standard Accessories:

- * Pulse oximetry sensors probe for adult blood pressure cuffs
- * Power cord with 3 prong tandem plug

Optional Parameters:

- * N20, O2, CO2 , IBP, Pacer rejection

Medical , Dental, Laboratory Equipment

APPROVED BUDGET FOR THE CONTRACT (ABC) : P 550,000.00

PEEDO-Davao del Norte Hospital IGACOS Zone

Period of Delivery : _____ day/s

**SUPPLIERS MUST SPECIFY/INDICATE
 BRAND NAMES UPON QUOTATION**

INSTRUCTIONS IN FILLING UP THIS BID FORM

Bids Distributed To:

1. Bidder must read carefully the statement at the back hereof.
2. Bid Quotation must be typewritten or printed in blue or black ink.
3. Unit price, total amount and grand/lot price must be properly indicated.
4. Delivery period must be indicated in the Bid form unless otherwise provided by the procuring unit/office. Failure to indicate the period of delivery shall automatically mean 10 days delivery after receipt of the Purchase Order.
5. Place of Delivery shall be at the PGSO Warehouse unless otherwise indicated in the bid form.
6. Submit Bid in the sealed envelope.
7. Unprinted name of establishment, unprinted name of the bidder and unsigned bid shall be disqualified.
8. Erasures/alterations should be initialed accordingly.

 Name of Establishment

 Name of Bidder

 Signature of Bidder

Republic of the Philippines
Province of Davao del Norte
Government Center, Mankilam, Tagum City

PR Number : 1301-0015

Bid Number : B20130004

Date : January 14, 2013

Page : 3

Item No.	Qty./Unit	Item	Bid Price	
			Unit Price	Total Amount

Optional Accessories: Mobile cart with lockable castors

Standard Requirements:

- * Manuals in English Language
- * Operation Manual and service Manual
- * Training Program for end-user on equipment's operation
- * One year warranty on parts and service after commissioning and acceptance
- * IEC /PNS compliant or its equivalent
- * Availability of Technician in Mindanao

* * * GRAND/LOT PRICE : P _____

Medical , Dental, Laboratory Equipment

APPROVED BUDGET FOR THE CONTRACT (ABC) : P 550,000.00

PEEDO-Davao del Norte Hospital IGACOS Zone

Period of Delivery : _____ day/s

**SUPPLIERS MUST SPECIFY/INDICATE
BRAND NAMES UPON QUOTATION**

INSTRUCTIONS IN FILLING UP THIS BID FORM

Bids Distributed To:

1. Bidder must read carefully the statement at the back hereof.
2. Bid Quotation must be typewritten or printed in blue or black ink.
3. Unit price, total amount and grand/lot price must be properly indicated.
4. Delivery period must be indicated in the Bid form unless otherwise provided by the procuring unit/office. Failure to indicate the period of delivery shall automatically mean 10 days delivery after receipt of the Purchase Order.
5. Place of Delivery shall be at the PGSO Warehouse unless otherwise indicated in the bid form.
6. Submit Bid in the sealed envelope.
7. Unprinted name of establishment, unprinted name of the bidder and unsigned bid shall be disqualified.
8. Erasures/alterations should be initialed accordingly.

.....
Name of Establishment

.....
Name of Bidder

.....
Signature of Bidder